

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

CLIENT: Box R Waterlab
Project: X022774 Heppner HS Lead Study
Lab Order: 1606B11

Date: 06-Jul-16

CASE NARRATIVE

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: **1606B11**
NRC Sample ID: **1606B11-01A**
Collection Date: **6/23/2016 9:57:00 AM**
Received Date: **6/27/2016 12:30:00 PM**
Reported Date: **7/6/2016 8:26:00 AM**

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #1
Collectors Name: SKM
Sample Location: W Entry Traditional DF Higher
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|--------|------|----------|-------|--------------|------------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | ND | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-02A
Collection Date: 6/23/2016 9:44:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #2
Collectors Name: SKM
Sample Location: South Gym DF
Source: City Water

ANALYTICAL RESULTS

| Analyses | NELAP | | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|--------|------|----------|-------|--------------|------------------|---------|
| | Method | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | ND | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-03A
Collection Date: 6/23/2016 9:40:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #3
Collectors Name: SKM
Sample Location: Cafeteria Veg Prep Sink
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|--------------|------------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00883 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-04A
Collection Date: 6/23/2016 9:00:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #4
Collectors Name: SKM
Sample Location: South Upper Hall DF
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|-----------|---------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00135 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-05A
Collection Date: 6/23/2016 9:43:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #5
Collectors Name: SKM
Sample Location: Gym Water Dispenser
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|--------|------|----------|-------|--------------|------------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | ND | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100018
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-06A
Collection Date: 6/23/2016 9:32:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #6
Collectors Name: SKM
Sample Location: Weight Room Balcony
Source: City Water

ANALYTICAL RESULTS

| Analyses | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|-------------------|------|----------|-------|-----------|---------------|---------|
| | Method | Accredited Result | | | | | | |
| Lead | EPA 200.8 | A 0.00182 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-07A
Collection Date: 6/23/2016 9:15:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #7
Collectors Name: SKM
Sample Location: Entry/Office DF
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|-----------|---------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00109 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-08A
Collection Date: 6/23/2016 9:42:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #8
Collectors Name: SKM
Sample Location: Boys Locker Room DF
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|----------|------|----------|-------|-----------|---------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.000742 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-09A
Collection Date: 6/23/2016 9:37:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #9
Collectors Name: SKM
Sample Location: EZ Dispenser West
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP Accredited | Result | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------------|--------|------|----------|-------|-----------|---------------|---------|
| Lead | EPA 200.8 | A | ND | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-10A
Collection Date: 6/23/2016 10:19:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #10
Collectors Name: SKM
Sample Location: Ag Building DF
Source: City Water

ANALYTICAL RESULTS

| Analyses | NELAP | | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|--------------|------------------|---------|
| | Method | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00179 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-11A
Collection Date: 6/23/2016 9:50:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #11
Collectors Name: SKM
Sample Location: Cafeteria Sink
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|--------------|------------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00442 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: **1606B11**
NRC Sample ID: **1606B11-12A**
Collection Date: **6/23/2016 9:17:00 AM**
Received Date: **6/27/2016 12:30:00 PM**
Reported Date: **7/6/2016 8:26:00 AM**

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #12
Collectors Name: SKM
Sample Location: Lower Hallway DF
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|--------------|------------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00127 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100018
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-13A
Collection Date: 6/23/2016 9:41:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #13
Collectors Name: SKM
Sample Location: Girls Locker Room DF
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|-----------|---------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00310 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-14A
Collection Date: 6/23/2016 9:36:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #14
Collectors Name: SKM
Sample Location: EZ Dispenser Gym North
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|--------|------|----------|-------|-----------|---------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | ND | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-15A
Collection Date: 6/23/2016 10:50:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #15
Collectors Name: SKM
Sample Location: West Entry Hallway Disp
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|--------|------|----------|-------|-----------|---------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | ND | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRI.

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606B11
NRC Sample ID: 1606B11-16A
Collection Date: 6/23/2016 10:02:00 AM
Received Date: 6/27/2016 12:30:00 PM
Reported Date: 7/6/2016 8:26:00 AM

Sample Information:

X022774 Heppner HS Lead Study

Client Sample ID: X022774 #16
Collectors Name: SKM
Sample Location: Staff Lounge
Source: City Water

ANALYTICAL RESULTS

| Analyses | Method | NELAP | | Qual | MRL | Units | EPA Limit | Date Analyzed | Analyst |
|----------|-----------|------------|---------|------|----------|-------|--------------|------------------|---------|
| | | Accredited | Result | | | | | | |
| Lead | EPA 200.8 | A | 0.00168 | | 0.000103 | mg/L | 0.020 AL | 7/3/2016 | OML |

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

Date: 06-Jul-16

CLIENT: Box R Waterlab
 Work Order: 1606B11

ANALYTICAL QC SUMMARY REPORT

Project: X022774 Heppner HS Lead Study

Batch ID: 35603

| | | | | | | | | | | | |
|------------|----------|-----------------|-----------------------|-------------|-------------------------|----------------|-----------|-------------|------|----------|------|
| Sample ID | MB-35603 | SampType: MBLK | TestCode: ICPMS_200.8 | Units: mg/L | Prep Date: 6/28/2016 | RunNo: 88186 | | | | | |
| Client ID: | ZZZZ | Batch ID: 35603 | TestNo: EPA 200.8 | (EPA 200.8) | Analysis Date: 7/3/2016 | SeqNo: 1318646 | | | | | |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | ND | 0.000103 | | | | | | | | | |

| | | | | | | | | | | | |
|------------|-----------|-----------------|-----------------------|-------------|-------------------------|----------------|-----------|-------------|------|----------|------|
| Sample ID | LCS-35603 | SampType: LCS | TestCode: ICPMS_200.8 | Units: mg/L | Prep Date: 6/28/2016 | RunNo: 88186 | | | | | |
| Client ID: | ZZZZ | Batch ID: 35603 | TestNo: EPA 200.8 | (EPA 200.8) | Analysis Date: 7/3/2016 | SeqNo: 1318647 | | | | | |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 0.09326 | 0.000104 | 0.1 | 0 | 93.3 | 85 | 115 | | | | |

| | | | | | | | | | | | |
|------------|---------------|-----------------|-----------------------|-------------|-------------------------|----------------|-----------|-------------|------|----------|------|
| Sample ID | 1606B11-12AMS | SampType: MS | TestCode: ICPMS_200.8 | Units: mg/L | Prep Date: 6/28/2016 | RunNo: 88186 | | | | | |
| Client ID: | X022774 #12 | Batch ID: 35603 | TestNo: EPA 200.8 | (EPA 200.8) | Analysis Date: 7/3/2016 | SeqNo: 1318670 | | | | | |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 0.09564 | 0.000104 | 0.1 | 0.001267 | 94.4 | 70 | 130 | | | | |

| | | | | | | | | | | | |
|------------|----------------|-----------------|-----------------------|-------------|-------------------------|----------------|-----------|-------------|------|----------|------|
| Sample ID | 1606B11-12AMSD | SampType: MSD | TestCode: ICPMS_200.8 | Units: mg/L | Prep Date: 6/28/2016 | RunNo: 88186 | | | | | |
| Client ID: | X022774 #12 | Batch ID: 35603 | TestNo: EPA 200.8 | (EPA 200.8) | Analysis Date: 7/3/2016 | SeqNo: 1318671 | | | | | |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 0.09452 | 0.000104 | 0.1 | 0.001267 | 93.2 | 70 | 130 | 0.09564 | 1.18 | 20 | |

Qualifiers: E Value above quantitation range
 ND Not Detected at the Minimum Reporting Limit
 H Holding times for preparation or analysis exceeded
 R RPD outside accepted recovery limits
 J Analyte detected below quantitation limits
 S Spike Recovery outside accepted recovery limits

Neilson Research Corporation

Date: 06-Jul-16

CLIENT: Box R Waterlab
 Work Order: 1606B11
 Project: X022774 Heppner HS Lead Study

ANALYTICAL QC SUMMARY REPORT

Batch ID: 35605

| | | | | | | | | | | | |
|------------|----------|-----------|-----------|-------------|-------------|-------------|-----------|----------------|-----------|----------|---------|
| Sample ID | MB-35605 | SampType: | MBLK | TestCode: | ICPMS_200.8 | Units: | mg/L | Prep Date: | 6/28/2016 | RunNo: | 88186 |
| Client ID: | ZZZZZ | Batch ID: | 35605 | TestNo: | EPA 200.8 | (EPA 200.8) | | Analysis Date: | 7/3/2016 | SeqNo: | 1318674 |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | ND | 0.000103 | | | | | | | | | |

| | | | | | | | | | | | |
|------------|-----------|-----------|-----------|-------------|-------------|-------------|-----------|----------------|-----------|----------|---------|
| Sample ID | LCS-35605 | SampType: | LCS | TestCode: | ICPMS_200.8 | Units: | mg/L | Prep Date: | 6/28/2016 | RunNo: | 88186 |
| Client ID: | ZZZZZ | Batch ID: | 35605 | TestNo: | EPA 200.8 | (EPA 200.8) | | Analysis Date: | 7/3/2016 | SeqNo: | 1318675 |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 0.09374 | 0.000104 | 0.1 | 0 | 93.7 | 85 | 115 | | | | |

| | | | | | | | | | | | |
|------------|---------------|-----------|-----------|-------------|-------------|-------------|-----------|----------------|-----------|----------|---------|
| Sample ID | 1606B13-16AMS | SampType: | MS | TestCode: | ICPMS_200.8 | Units: | mg/L | Prep Date: | 6/28/2016 | RunNo: | 88186 |
| Client ID: | ZZZZZ | Batch ID: | 35605 | TestNo: | EPA 200.8 | (EPA 200.8) | | Analysis Date: | 7/3/2016 | SeqNo: | 1318698 |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 0.1101 | 0.000104 | 0.1 | 0.01646 | 93.7 | 70 | 130 | | | | |

| | | | | | | | | | | | |
|------------|----------------|-----------|-----------|-------------|-------------|-------------|-----------|----------------|-----------|----------|---------|
| Sample ID | 1606B13-16AMSD | SampType: | MSD | TestCode: | ICPMS_200.8 | Units: | mg/L | Prep Date: | 6/28/2016 | RunNo: | 88186 |
| Client ID: | ZZZZZ | Batch ID: | 35605 | TestNo: | EPA 200.8 | (EPA 200.8) | | Analysis Date: | 7/3/2016 | SeqNo: | 1318699 |
| Analyte | Result | MRL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | RPD Ref Val | %RPD | RPDLimit | Qual |
| Lead | 0.1094 | 0.000104 | 0.1 | 0.01646 | 92.9 | 70 | 130 | 0.1101 | 0.663 | 20 | |

Qualifiers: E Value above quantitation range
 ND Not Detected at the Minimum Reporting Limit

H Holding times for preparation or analysis exceeded
 R RPD outside accepted recovery limits

J Analyte detected below quantitation limits
 S Spike Recovery outside accepted recovery limits